

UNCLASSIFIED//~~FOR OFFICIAL USE ONLY~~



Defense Intelligence Reference Document

Acquisition Threat Support

23 March 2010

ICOD: 1 December 2009

DIA-08-1003-012

Technological Approaches to Controlling External Devices in the Absence of Limb-Operated Interfaces

UNCLASSIFIED//~~FOR OFFICIAL USE ONLY~~

Technological Approaches to Controlling External Devices in the Absence of Limb-Operated Interfaces

Prepared by:

(b)(3):10 USC 424

Defense Intelligence Agency

Author:

(b)(6)

Administrative Note

COPYRIGHT WARNING: Further dissemination of the photographs in this publication is not authorized.

This product is one in a series of advanced technology reports produced in FY 2009 under the Defense Intelligence Agency, (b)(3):10 USC 424 Advanced Aerospace Weapon System Applications (AAWSA) Program. Comments or questions pertaining to this document should be addressed to (b)(3):10 USC 424;(b)(6), AAWSA Program Manager, Defense Intelligence Agency, ATTN: (b)(3):10 USC 424 Bldg 6000, Washington, DC 20340-5100.

Contents

Introduction.....	v
Direct Neural Signals.....	1
Indirect Neuronal Signals – The BDL D Effect.....	3
Control of External Devices	4
Noninvasive Technologies	6
EEG	7
MEG.....	8
EMG.....	9
MRI and fMRI	9
NIRS.....	10
Invasive Technologies.....	10
Open-Loop Direct Cortical Array Algorithm Modeling	11
Closed-Loop Peripheral Arrays Utilizing Visual Feedback	13
MEMS, ECoG, and PSoC Circuitry	14
Chronic Neural Implants and fMRI	16
New Electrode Designs.....	18
Hybrid Neuro-Robotic Systems for True Closed-Loop BMI.....	20
Trials Using Human Subjects.....	22
Optical Stimulation of Action Potentials	23
Discussion	23
Noninvasive Electrical Devices	24
Noninvasive BDL D-Based devices.....	25
Noninvasive Magnetic Devices	26
Invasive Technologies – General, Optical, and Ex-Vivo Engineering.....	26
Implantable Chips, Ladders, and Arrays.....	27

Conclusions 27

Figures

Figure 1. Simplified Rendering of a Neuron 1
Figure 2. The General Layout of a Closed-Loop Control Interface..... 5
Figure 3. Two Commercially Available EEG Sensors..... 8
Figure 4. Experimental Overview of Brain-Controlled Robot in a Closed Loop
With Visual Feedback Experiment 13
Figure 5. Schematic of the Neurochip Functional Blocks 15
Figure 6. Implant Location 16
Figure 7. Histology and Electrode Tracks 17
Figure 8. The Michigan Electrodes 18
Figure 9. Image Distortion and Custom Microwire Electrode Assembly to
Improve It..... 19
Figure 10. Example of T2 Variability..... 19
Figure 11. T2 Value Analysis Summary of T2 Values in All the Image Slices That
Spanned the Electrode Arrays in All Animals 20
Figure 12. A Hybrid Neuro-Robotic System..... 21
Figure 13. Experimental Arrangements 21

Technological Approaches to Controlling External Devices in the Absence of Limb-Operated Interfaces

Introduction

Since the advent of modern interactive control of computers, technologies have been sought to directly connect the biological and the physical to form a seamless entity. While science fiction explores the possibilities of shared consciousness between brains and mainframes, real-life scientists work toward an equally fantastic but more pedestrian goal of eliminating required electromechanical human-machine interfaces (HMIs). Such technologies promise integration of thought and computer-controlled action, without the need for a limb-operated device to translate intent from physiological networks to physical circuits.

This paper first briefly reviews underlying neural structure, function, and activity to provide background on the type of signals and the scale of temporal changes that arise from conscious control within the nervous system. A short primer on brain-machine interfaces (BMIs) is included at the end of the background material. This is followed by a survey of the state-of-the-art detection and stimulation technologies available utilizing noninvasive and invasive methods, including several studies that exemplify research paths currently being undertaken. Discussion brings together elements of the technology survey, application limitations, a timeline for useful commercial deployment, and predicted future research directions. Two technology research paths are highlighted as the most probable to produce functionally useful devices (defined as nearly equal or even superior to the control capabilities of current HMIs) in the near and far term.

The technology endpoint this paper seeks is thought-based operation of remote machinery during normal human activities without mechanical device interaction—that is, control of external devices without the need to go to a specified location, such as a shielded room; without the need to remain perfectly motionless to reduce signal noise; and without the need for interaction with a normal electro-mechanical device, such as an i-Phone or other handheld device with buttons and trackballs. Data transfer rates are sought to exceed 5-10 bits/second to be useful for operation of complex devices; this rate range and above is referred to as high-bandwidth BMIs. Response time for a 1-of-N selection of commands is targeted at 300 milliseconds or less. These are the operational parameters assumed for final implementation unless some limitation or supercapability is described.

The paper concludes that noninvasive electrical monitoring of neural activity, primarily reading combined action of muscles and neurons, is the most promising commercial technology in the near term. Inherent limitations in the noninvasive electrical approach require future development of different research paths. The paper argues, mainly by eliminating other approaches, that the most probable technology in the long term involves invasive single-neuron-based direct cortical connections to form a network of high-bandwidth duplex communication pathways. Currently the most promising technologies

for robust construction of such an interface are optical stimulation, gating, and sensory devices, and chip-based electrode arrays that have been encased in ex-vivo-engineered neural tissue.

Direct Neural Signals

The human nervous system has two classes of cells, neurons and glia. Based on all research to date, it is believed that signals within the network of neurons constitute the whole of information processing, with glial cells playing a purely supporting role. This neural doctrine has dominated research in BMI until recently and still constitutes the only major research path in direct technologies. Furthermore, all technologies directly measuring human neuronal action rely on detecting or influencing electrical activity of these cells; no current in situ research selectively affects neurotransmitter activity between local cells for the purpose of information exchange. Therefore, the focus for the foreseeable future will be on the electrical activity of neurons as the primary target of BMI.

Neurons consist of four parts: axon, dendrites, cell body or soma, and pre-synaptic terminals. Electrical information is transmitted to the neuron through the dendrites, proceeds through the cell body, and leaves the cell through the axon at one or more pre-synaptic terminals. Neurons have one axon and from one to tens of thousands of dendrites.




Figure 1. Simplified Rendering of a Neuron. The arrows indicate the direction in which signals are conveyed. The single axon conducts signals away from the cell body, while the multiple dendrites receive signals from the axons of other neurons. The nerve terminals end on the dendrites or cell body of other neurons or on other cell types, such as muscle or gland cells. (Reference 1)

Chemical details of how the action potentials travel through the cell or are transmitted across the synapse are not important to the current treatise, other than the distinction that in these biologically based electrical networks, ions of sodium, potassium, and chlorine move through the cell membranes perpendicular to the propagation of the action potential down the axon. This allows information to be transmitted faster than ions could flow down the axon. The propagation of information is similar to a wave traveling down a garden hose: quickly move one end of the hose back and forth with sufficient force, and a wave will travel to the other end of the hose; however, any part of the hose structure has only moved (nominally) perpendicular to the direction of wave propagation. In a similar fashion, ions flow through channels across the axon's cell membrane, changing the local membrane potential and thus propagating the electrical signal down the axon.

The signal transmission down the axon of a neuron is an all-or-nothing process. When the cell body is stimulated above threshold, the axon transmits the same action potential at the same speed and in the same direction, regardless of the extent above threshold or duration of the input. Action potentials have durations of 1-10 milliseconds. Input signals can result in transmission of multiple action potentials, and thus the frequency and number of neuronal firings do vary with the input. Neurons require some time to reset between firings, nominally the duration of the pulse for that axon, yielding a typical maximum firing rate of between 100 hertz (Hz) and 1 kilohertz (kHz).

It is instructional at this point to contrast this mechanism with propagation of signals through a physical electrical circuit, the planned external portion of our BMI. In a copper wire, electrons carry the signal. Electrons drift along the signal path, but the signal itself moves as a compression wave rather than a transverse wave as in the biological system. Going back to our garden hose example, consider the hose now filled with small marbles: inserting a marble at one end will move each marble in the hose just a little, but very rapidly the last marble in line will pop out of the far end of the hose. As with the biological system, the signal is propagated to the far end of the hose by local actors rather than physical motion of a single ion or electron moving the whole distance. The underlying physics governing the signal transmission makes metallic and semimetallic circuits about a million times faster than the biological system.

This difference in the carriers and underlying mechanisms of signal transmission between biological and physical circuits has so far prevented the invention of a direct connection between the two disparate systems. Instead, both noninvasive and invasive direct-detection technologies rely on placing physical sensors or transmitters in close proximity to the neurons of interest and utilizing classical electrodynamics to govern signal jump between the systems. Additionally, given the maximum typical firing rate for neurons of 1 kHz, sampling of action potentials at a few kHz will be fast enough to detect firing of any individual neuron, though super-sampling above 10 kHz can be used to reduce noise. Higher sampling frequencies also may be required if multiple neurons are monitored and quantitative information about their relative firing sequence is desired. Frequencies of a 1 kHz or below are sufficient to stimulate action potentials, and again, higher system frequency may be required for multiple neuron sequential stimulation. Finally, higher frequencies may be required if monitoring or stimulation of some aspect of signal transmission other than action potentials is sought, such as monitoring single ion channels.¹ There is currently no evidence that transduction frequencies above 100 kHz have any advantage in BMIs, thus the main challenge is the connection dynamics, since even this ultra-maximum frequency is easily attainable with current electronic manufacturing technology.

The human brain doesn't process information as a traditional computer does. Information is moved around through pathways and at certain neurons it is allowed or not allowed to pass based on excitatory or inhibitory dendritic signals arriving before triggering of action potentials. Local groups of neurons can act nearly coherently, for example in volition of motor action like a hand movement. Detecting such coherent firing at nodes around the brain is robust both noninvasively and invasively, though

¹ Technologies that monitor single ion channels on a neuronal membrane are important for research on neural function and neurotransmitter action. However, these devices are ultra-sensitive to physical motion, and have so far not proven useful in large scale information transfer required in BMIs.

noninvasive techniques currently cannot resolve firing sequences of individual neurons within such groups.

Living neurons in an active tissue fire in the rest state. Changes in the frequency of these firings imply that a given neuron is currently involved in the processing of information. Bulk changes in local field potential oscillations imply that several neurons are active. This is the signal seen in the noninvasive direct-measurement techniques of electroencephalography (EEG) and magnetoencephalography (MEG).

Movement as simple as an eye blink involves signal communication through a million neurons. Several locales in the nervous system thus offer the possibility of placement of an invasive monitoring device for detecting electrical activity associated with an eye blink or other signal of interest. Many other factors will influence the placement decision, but considering only catching the signal as an action potential, using any neuron along the complete pathway is equally as good, assuming detection of individual neuronal firing can be accomplished.

Considering cortical placement of detection devices, the most common area of activity under study is the outer covering of gray matter of the cerebral cortex, the neocortex. The neocortex is about 2.5 mm thick in humans and follows the ridges and fissures of the brain (gyri and sulci, respectively, or gyrus and sulcus, if singular). The neocortex is roughly divided into six layers and different cortical probes can concentrate on activity in different layers, or just measure the combined activity on the surface of the cortex, or the combined activity that is detectable noninvasively through the layers of tissue and bone of the cranium.

Detecting single firings of individual neurons is a difficult process because the signals are weak to start with, and not isolated from the rest of the electrical activity within the brain. Large groups of coherent neurons, perhaps a few thousand to tens of thousands all firing at once in relation to an external event, are the most studied of single firing signals.² One of the most studied signals is the so-called P300 which occurs in many cognitive tasks (References 2-4).

P300 is a term used to indicate a pulse signal from a large group of neurons that appears about 300 milliseconds after a stimulus event. Peaks that occur between 200 and 400 milliseconds are commonly grouped into the P300 category. The "P" stands for a positive measured voltage, while "N" signals are negative.³ The P300 signal has been shown to be influenced by top-down executive function and is therefore a prime candidate for a trainable interface (Reference 5). Earlier signals, like P50 and N100, are of a greater interest for differential diagnosis of pathology in current research.

The preceding discussion is greatly simplified version of the electrical dynamics of neuronal firing and does not include differences between axon and dendrite signals, or the transmission of signals across the synapse. Complete discussion of underlying electrical signals in the nervous system can be found in (References 6, 7).

Indirect Neuronal Signals – The BOLD Effect

² Single firing signals here mean a single peak of combined electrical activity relative to an event. This is not necessarily the same as single firings of each neuron.

³ The literature is not consistent with plotting P signals up, and N signals down; however, P peaks are always in the opposite direction of N peaks.

The brain activity mentioned above is a complex chain of ionic motion within the central nervous system. Ionic production, release, and movement within and between cells all consume energy. This energy is supplied by the conversion of blood-borne oxyhemoglobin to deoxyhemoglobin. The rate of oxygen consumption in a localized volume varies based on local neural activity. The circulatory system compensates for changes in energy demand by increasing or decreasing both the flow rate and volume of blood, regionally and locally. Local energy demand, expressed in the capillary beds, will alter the rate at which oxygen is metabolized, called the cerebral rate of oxygen metabolism, abbreviated CMRO₂. When brain activity increases in a region, the circulatory response, called the hemodynamic response, will be increases in flow and volume, while the local areas increase CMRO₂.

The hemodynamic response consistently provides an excess of oxygen over what is required, and this results in some oxyhemoglobin traveling through the capillary bed and local venous structure without being converted into deoxyhemoglobin. Oxy- and deoxyhemoglobin have different magnetic susceptibilities, and different infrared spectra. The hemodynamic response, by changing the net ratio of oxy- to deoxyhemoglobin in the local venous structure, thus changes the local magnetic susceptibility and local infrared resonance spectra around focused brain activity. This complex chain reaction is called the Blood Oxygen Level Dependent, or BOLD effect (Reference 8). The BOLD effect leads to a method to indirectly measure the local brain activity by monitoring the hemodynamic response using Magnetic Resonance Imaging (MRI) or Near Infrared Spectroscopy (NIRS).

The BOLD response to any event peaks about 4-6 seconds after the event occurs, limiting the applications for which monitoring these signals and their associated delay may be useful, given the BMI operating parameters defined above. Furthermore, person to person variation in distributed signals show significant differences in regions activated (Reference 9), though there is evidence that these inter-subject variations are stable intra-subject over time (Reference 10).

Control of External Devices

In general, there are two types of BMI design called open-loop and closed-loop. Open loop designs, those either detecting activity or providing stimulation, are important for BMI research and application other than the practical control of an external device. Closed-loop systems include both output signals from the brain to effect a change in the state of a device, and a stimulation channel, which can be used as feedback. In some system designs, the feedback can be via visual cues such as watching a display (Reference 11). Open-loop systems are important for BMI research and have application in sense replacement such as artificial hearing, or when extensive output signal processing is combined with external device state feedback (see remote robot operation below). For learning and fine, efficient control of external devices without significant external processing, closed-loop systems are typically needed.

Several components constitute a general BMI system for control of an external device (Figure 2). As mentioned in the previous section, the interfaces where the information signal makes the jump between biological and physical pathways are the most challenging part of the system. These connections can be placed to stimulate or read peripheral or cerebral neurons. The processing step could be as simple as amplification

of a spiking train to trigger a switch, or as complex as decoding signal from noise utilizing a 300-channel EEG cap.




Figure 2. The General Layout of a Closed-Loop Control Interface. The input and output interfaces can attach to read or affect activity of cerebral or peripheral neurons. Additional sensory feedback channels such as visual displays, audio tones, or haptic feedback are possible. In some system designs, especially noninvasive types, these sensory feedback channels replace stimulation inputs. In an open-loop interface, either the input or output signal paths are absent.

One consideration in developing a BMI control system is the utility of the application: does it make sense to directly tie these controls to neural activity? The bulk of research in direct neural interfaces is toward the goal of restoring mechanical capabilities to those individuals who have either lost limbs or lost control of limbs through central nervous system injury or disease. In these cases it is of obvious utility to produce a system that moves a cursor across screen to select an item over the course of several seconds; however, for a healthy individual, it is far more practical to execute a hand movement to control a mechanical device⁴ to achieve the same goal.

An exception to consider is the case of applications where the subject's hands or feet are already occupied with other essential tasks in the overall application. A research question for any such application is whether training a subject to utilize a BMI is advantageous over a more complex mechanical interface. Furthermore, whether training on a new BMI affects learning/retaining proficiency with similar BMIs is an additional concern.

Moreover, for any BMI, one can quantitatively describe the information transfer in controlling the magnitude of human motor action (Reference 12). Since the advent of digital control of external devices, this information is expressed in equivalent bits per

⁴ Mechanical device as used here represents any limb-operated interface, such as a mouse or touchscreen.

second transferred. Higher bit rates equate to more efficient interfaces. Baseline examples of performance include finger pointing, which can convey 14 bits/s of information (Reference 13), operation of a mouse about 8 bits/s (Reference 14), while stylus tapping a soft QWERTY keyboard on a PDA has a lower rate, around 5 bits/s (Reference 15).

Once these questions are researched and answered for a given application, the design choice should be made whether a new mechanical interface, or an optical interface that tracking detailed motion of the eyes, body and limbs, may be the optimum solution rather than a BMI.⁵

Another consideration in developing a BMI control system is the cognitive limit of supervisory control of external devices. The current state of the art for human interfaces involves volition of control, meaning that there is some executive supervision, typically involving the pre-frontal cortex, active during the control process. Recent studies relating to control of semiautonomous vehicles show that there are significant limits on the number of supervisory activities that can be undertaken simultaneously (Reference 16). In response to these limits, one can consider utilization of nonsupervised brain networks in control of external devices, similar to those brain systems that initiate very complex muscle coordination activity such as martial arts or playing a musical instrument, or systems like those that are responsible for autonomic function. Application from such research is long term, but there are current studies that show BMI training of at least one spinal cord patient exhibited differential activity in the cerebellum, owing to extraordinary plasticity (Reference 17). It is therefore not out of the realm of possibility that sufficient input and output interfaces that include neural pathways through the cerebellum could result in over-learning a BMI application to the point that supervisory control is developed and thus significant cognitive effort is not required.

Noninvasive Technologies

There are four noninvasive technologies currently in widespread use in monitoring brain activity. The two direct-measure technologies involve detecting surface currents from the relative voltage changes at or just below the scalp, electroencephalography (EEG), and detection of near scalp magnetic fields associated with neural pathway current flow, magnetoencephalography (MEG). The remaining two indirect measures involve monitoring BOLD response through rapid successions of whole brain MRI (functional MRI or fMRI), and near-infrared spectroscopy (NIRS).⁶

The primary technology used for BMI is also the oldest. EEG was first described in 1929 (Reference 18) and now exists in several derivative forms. Traditional EEG uses electrodes at the surface of the scalp to measure and amplify potential differences between points above the cortical surface and a fixed reference such as the average reading from the earlobes. Traditional EEG data is analyzed by breaking up the power spectrum into several bands between 0.5 and 100 Hz. A derivative form of EEG developed about 10 years later is called event-related potentials, or ERP. In ERP, scalp

⁵ There is a valid argument to be made that optical tracking and analysis, especially of the eyes, is a BMI technology as defined in the current treatise. It is excluded as one merely by fiat.

⁶ NIRS is occasionally referred to as functional NIRS or fNIRS. NIRS using multiple sources to produce 3D images of internal changes in blood flow is occasionally called diffuse optical tomography (DOT). DOT is a more general technology that can also refer methods such as to visible-light laser excitation of tissue.

data is averaged over several electrodes time-locked to a stimulus (Reference 19). Both the previous methods record summed electrical activity of nominally 50,000 local neurons, thus large coherent group spiking activity⁷ is required to produce appreciable signal.

EEG

EEG-based BMI systems use pattern recognition among the several electrodes to transmit information. In closed-loop systems, audio or visual feedback is utilized. Training on the system and adaptation by the processing software can increase target detection efficiency, though some recent work shows that systems can be produced that nearly work right out of the box. Blankertz demonstrated a system that required only 20 minutes of training on naïve subjects (Reference 20).

Development is partially being driven by the need to provide a means of communication or action on the environment to patients that have lost control of their body. Much of this clinically-orientated research has focused on 'locked-in' patients, who suffer from total paralysis following brainstem stroke or degenerative diseases such as amyotrophic lateral sclerosis (ALS) (Reference 21). The goal has been to extract control signals either from surface EEG signals or from electrodes implanted near or within the cerebral cortex: ECoG (see next section on invasive technologies for details on ECoG). ALS deteriorates peripheral motor function before progression to cognitive areas. This greatly reduces a primary source of bioelectric noise, cranial and facial muscle action, and thus provides a reduced signal processing problem to decode neural signals. Successful communication has been established via EEG BMI in several studies based on both spiking activity and P300 signals (References 22-25). The response time to execute a command using these systems is measured in seconds. The results from ALS patients represent a best-case scenario for what could be accomplished using EEG-only sensors in a normal, healthy human who would have significant muscle noise to sift through.

The number of electrodes and the time-consuming application of conducting gel make EEG arrays bulky and slow to put on. An improvement would be a reduction in the number of electrodes, typically a few dozen for traditional EEG, and electrodes that could operate without conducting gel. Commercial versions of dry-electrode devices are being released on the market from companies such as Neurosky, OCZ Technology, and Emotiv.⁸ The Neurosky and Emotive systems claim proprietary processing algorithms, lack peer-reviewed literature supporting their claims, and even discuss using facial muscle movement to send signals (References 26, 27), raising doubt about whether neural signals are being measured at all. The NIA, for Neural Impulse Actuator, from OCZ clearly states that the electrodes pick up a combination of EEG, EMG, and EOG,⁹ and that the algorithm is only interpreting the net signal patterns instead of trying to sort out the EEG component.

⁷ Spiking activity is the term used for recognition of action potentials. Spikes are fast and easy to recognize with electronic triggering circuits, while more complex waveforms require additional processing.


⁸ Websites www.emotiv.com, www.ocztechnology.com, and www.neurosky.com, last accessed 12 May 2009.

⁹ The electro-myogram (EMG) reads signals from muscle activity, in this case facial muscles. The electro-oculogram (EOG) measures electrical signals arising from muscles associated with the eyes.

Commercial EEG sensors



MindSet, by NeuroSky



nia, by OCZ Technology

Figure 3. Two Commercially Available EEG Sensors. The MindSet utilizes two sensors, one in the right earphone and one for placement on the forehead (what looks like a microphone arm is to be pressed against the forehead). The nia records signals from three sensors on the forehead.

In a documented and peer-reviewed study, Popescu and colleagues have shown a system that uses 6 dry electrodes and is 90 percent accurate in operation of a 1-D cursor by untrained subjects. Additionally, a measurement of error rates as a function of number of electrodes show that an increase to 12 electrodes could drop the error rate to around 5 percent, but that additional electrodes much beyond a dozen do not significantly improve the accuracy of their algorithm (Reference 28).

MEG

MEG-based BMI systems have been shown to be feasible utilizing a subject imagining limb movements for binary decisions (References 29, 30). Future work could improve the methodology to parallel that achieved with EEG; however, there are significant technological hurdles to a field deployment of MEG. The magnetic fields from neural activity are detected with very sensitive devices called superconducting quantum interference devices, or SQUIDs. These detectors are sensitive to the neural activity induced 100 fT (femtoTesla) changes near the scalp. SQUIDs only operate at very low temperatures thus requiring a cryogenic system as well significant detection and amplification electronics. The availability of a cold sink and high vacuum, such as in a space-based application, could reduce the support system overhead, but there is still the matter of the weakness of the signals. Even if future detector development like atomic magnetometers would solve the equipment overhead issue for Earthbound application, the fact that 100 fT is about 100 million times smaller than the Earth's magnetic field will prove an insurmountable barrier to sifting signal from noise in anything but a heavily-shielded, metal-free environment. Attempts have been made to use High- T_c superconductors as a shielding material with some success (Reference 31), but significant further development is needed.

EMG

Instead of measuring nerve impulses and amplifying them, one can design a system that monitors muscle movement in a hands-free approach to a mechanical interface, essentially using the muscles as a biological amplifier of neural signals. One may close the feedback loop using a heads-up display, monocle, or other visual device, or another traditional sensory feedback. The advantage of such a design is that controls may be actuated through a lightweight wireless system, effectively providing control stations anywhere such a wireless system would work. Such a system could be used to untie a pilot from the cockpit. Recent work has shown that performance of 1-2 bits/s is possible with minimal training, about four times the current performance of a comparison EEG forehead sensor (Reference 14).

MRI AND FMRI

Magnetic Resonance Imaging (MRI) works by a simple excitation and relaxation of spin states. When molecules containing hydrogen are placed in a strong static magnetic field, a small but detectable number of hydrogen protons align their intrinsic spins along the direction of the external field. An applied radiofrequency (RF) pulse near the proton resonant frequency, 42.6 MHz/Tesla or 128 MHz at 3 Tesla, knocks the spins perpendicular to the field and the relaxation back to ground state releases RF energy in patterns that can be reconstructed to show both composition and distribution of any hydrogen-rich material. The resonant frequency is a direct function of the local magnetic field defined by the Larmor relation: $\omega = \gamma B$; where ω is the frequency of precession, B is the local magnetic field, and γ is a constant of the material, 42.6 MHz/Tesla for bare protons as mentioned above.

Small perturbations to the static field will change the resonant frequency. By applying a small gradient to the static field, for example 100 milliTesla/meter along the z-axis, and limiting the bandwidth of the RF excitation signal to $\delta\omega$, one may select a slice of the brain perpendicular to the z-axis for excitation to δz . A change in the gradient field will change the position of the excited slice for the next excitation. Similar gradients in the x and y directions can limit the excitation to a single small volume of brain tissue. In current MRIs, these gradient fields are produced with electromagnets, and the series of time-dependent imaging gradient manipulations is called the scan sequence.

Free hydrogen (H) would produce a resonant signal slightly different than the bare proton due to the local field changes induced by its valence electron. Hydrogen gas (H₂) would produce a still different frequency since the local field around each proton is altered by the two shared electrons. Water molecules (H₂O) contain two hydrogen atoms and an entirely different "electron shield" than either H or H₂ and thus shows still another slightly different resonant frequency. Fat and other lipid molecules, important cell structure building blocks, have long chains of hydrocarbons, and the resulting ensemble of electron screening produces a wide peak that is substantially shifted¹⁰ from that of water.

Brain gray matter and white matter have different macroscopic lipid content and are thus able to be differentiated in an MRI scan. Different signals also arise in bone,

¹⁰ Frequency detection sensitivities in MRI are very good, and "substantial" here means about 3 parts per million. The frequency shifts caused by imaging gradients ranges in the parts per thousand.

cerebral-spinal fluid (CSF), and internal tissue structures of various other organs. Unlike x-ray based technologies, MRI scans can be optimized to contrast any of the many parts of the physics signal: total density of protons, water content, lipid content, and even particle motion in advanced techniques involving diffusion or spin labeling. Using such scans sequences that take several minutes, one can construct very high resolution images of gray and white matter structure for comparison with, and also mapping onto a "standard brain" template.

In addition to electron screening, macroscopic susceptibility will also change the local response to RF stimulation. The presence of even a small amount of metal, say as small as a hairpin, will greatly distort the reconstructed images. Indeed, the usual effect is a shift of frequency completely outside the sensitivity of the machine RF receiver in what is known as "drop-out." Smaller changes in local susceptibility, like produced in the BOLD effect, are measurable.

A series of fast scan sequences, typically collecting an entire brain volume at a resolution of 3 mm³ in 2 seconds, that are calibrated to optimize detection of the BOLD signal will show the dynamics of brain function under the specific internal or applied conditions; this is known as a functional MRI, or simply fMRI.¹¹ The major advantages of fMRI are unmatched 3D spatial resolution, compared to other noninvasive imaging methods, and complete skull penetration, making it the only imaging modality to unambiguously detect limbic activations important for determining emotionally-laden neuropsychological states. The main disadvantage for BMI is that the BOLD signal is detected several seconds after the neuronal firing takes place, making fMRI inappropriate for many naturalistic applications.

A long term prospect, likely in the 20- to 40-year timeframe, is that combined low-field MRI and MEG technology could detect neuronal firing deep in the brain and with high temporal accuracy. Initial experiments indicate some level of feasibility, but there is substantial development work required in room temperature low field magnetic field detection devices, such as atomic magnetometers, and signal processing algorithms to sift through the substantial electromagnetic background (References 32, 33).

NIRS

Near-infrared spectroscopy is an additional technology to monitor the BOLD effect noninvasively. Studies have shown it correlates well with the fMRI signal in animal models, although with reduced coverage and lower resolution (Reference 34). This lowered resolution greatly affects the accuracy of the technique, with recent work involving single trials and a decision attaining only 80 percent accuracy (Reference 35).

Invasive Technologies

The most prolific invasive BMI for use in humans is the cochlear implant (Reference 36), a sensory neuroprosthesis designed to aid in hearing for deaf individuals. This device, under continued development and refinement for more than 30 years, consists of a microphone, sound processor, and a receiver that is attached to an array of

¹¹ Specifically this is T2* Echo-Planar Imaging, also called BOLD EPI, Gradient Echo EPI, or BOLD fMRI. This approach is used in well over 95 percent of published functional studies, though there are more advanced techniques that concentrate on smaller portions of the hemodynamic signal. For example, Spin-Echo EPI will provide a higher localization within the gray matter, but the cost is a loss of 90 percent of the signal amplitude.

currently 22 electrodes directly implanted into the cochlea of the inner ear. The microphone and sound processor replaces the outer and middle ear, while the electrodes replace the frequency selective hair fibers of the inner ear that normally would transmit electrical signals to the cochlea. The quality of the sound is much less than natural hearing, but the only option for patients who have lost hearing or were born deaf. It has been estimated that 100,000 cochlear implants provide sound to users worldwide (Reference 37). The cochlear implant is an open loop stimulation system that provides complex signals to the nervous system through a peripheral connection.

BMI technologies have evolved from assistive technology devices solely targeted for the healthcare industry, into apparatuses designed for intracortical microstimulation to deliver sensory feedback. Such systems allow the simultaneous recording and microstimulation of neuronal and behavioral events. The primary device that has emerged leading this technological transformation is the direct cortical array containing either surface or neocortex penetrating electrodes. The direct cortical array is implanted on the surface of the brain near the functional area of interest, and then the connection is reverse-engineered: individual electrodes are tested to determine if they should be assigned as input, output, or neither.¹² These devices are designed to fuse neural signaling pathways with external machinery into a 3D control system, such as those apparatuses designed to modulate the movement of appendages (e.g. robotic arms), though the signals could be used for control of any complex machine. The control circuits on these systems are intricate and often bulky, and they can be designed with either open- or closed-loop sensory feedback.

An additional device is the penetrating cortical electrode. These electrodes contain a ladder of input/output probes and can be selected to transmit or receive signals in any layer of the neocortex. Research is ongoing regarding the optimal layer with which to place an interface. When using penetrating electrodes, tissue response is an additional challenge as these devices frequently cause scarring. Animal studies are in progress to quantify this effect.

Technological challenges in development of all types of invasive BMI devices include signal decoding/stimulation algorithms, and localization of brain activation near the implant. Functional MRI (fMRI) is utilized in conjunction with the invasive implants to confirm additional neural network activity during interface tasks. Modifications in the design of electrodes is necessary for experiments involving fMRI (Reference 38).¹³

The majority of invasive experiments to date use animal or even tissue models, though a few human trials have been conducted (with the exception of the cochlear implant described above).

OPEN-LOOP DIRECT CORTICAL ARRAY ALGORITHM MODELING

In traditional open loop experiments, eye movement systems are utilized to develop a linear model for a physiological system.¹⁴ In such experiments, invasive signals are recorded from the cortex and the motion of the eye is recorded in tandem. The location

¹² This procedure eliminates the need to directly locate the neuron(s) of interest during implantation.

¹³ Field potentials at the points of metal electrodes cause concentration of RF energy from the excitation pulses of the MRI. This effect is used in a similar setup for tissue ablation therapy; however, such energy concentration is undesirable in the studies at hand.

¹⁴ These techniques may be generally applied to any motor control systems.

on the cortex to implant cortical arrays is determined from previous experiments in fMRI (References 39, 40). In this way, a map is constructed correlating electrical neural activity with muscle movement. The goal of this exercise is to develop an algorithm that will predict which muscles move based on reading the neural activity alone. The reading of the neural activity for eye movement can then be used to move a device such as a camera lens.

There are two main models of fine motor control, one where the cortical motor areas perform all of the control functions and receive all sensory feedback, and a second where the cortical areas direct the function and receive interpreted feedback through sub-cortical or even peripheral networks. For open-loop invasive BMI applications, this is an academic question since peripheral interfaces would receive and send the same signals in both models, and cortical interfaces would blindly adapt external decoding algorithms based on the signals present regardless of model.

There have been numerous demonstrations of nonhuman primates controlling robots or graphical cursors in real-time through signals collected from cortical areas that employ open loop experimentation (References 41, 42). Kim, et al. conducted experiments where monkeys are trained on tasks prior to implantation, and then the tasks are repeated multiple times while muscle action and cortical activity are monitored (Reference 42). In these trials, shoulder and elbow torque were measured while the arm itself was constrained in an exoskeleton such that the hand would only move in a plane axial to the monkey's torso. A visual cursor was introduced and projected on a screen above the monkeys hand to follow the 2-D motion from the center starting point to the various task targets, which are also projected on the screen. The shoulder and elbow position recorded the state of flexation of 6 sets of muscle groups, collectively called the musculoskeletal arm model (MAM). Relating the spiking activity from implanted arrays to even this simplified 2-D MAM motion proved quite complex, and no fit correlating the observed movements and neural activity could be obtained with a linear model when kinematic impedance was considered.¹⁵

Even considering the six inputs, the 2-D problem is essentially a computer cursor control and therefore a relatively simple device, fully specified by a Cartesian coordinate system. The ultimate goal of these control systems is to manipulate something much more complex, like an arm, which may have many more degrees of freedom organized in a completely different coordinate system. For these tests a more elaborate "tracking system" may be utilized in teaching a primate to feed itself¹⁶ using a directly observed cortically controlled robotic arm.

Open-loop control systems have an inherent drawback in cases where cortical activity controls movement directly via an adaptive algorithm. Training the algorithm is the critical part of interface development. The adaptive algorithms use an iterative process to create a brain-to-cursor motion decoding scheme based on how the neurons fire when different targets are presented and therefore rely on previous normal feedback training – the brain knows how to move an arm because it has been moving an arm for most of its life.

¹⁵ Impedance to motion is essential for realistic operation of artificial limbs.

¹⁶ Food in this experiment is used as a reward.

An issue develops when an entirely new device is considered. The form of the movement control algorithm is similar to a population vector in that movement at each time step is determined by a vector sum of the neurons' normalized firing rates multiplied by a set of linear coefficients. Modern neuroscience has proven that tens of thousands, even up to millions of neurons fire in concert to perform even the most mundane of movements. These extended neural pathways developed through full-duplex closed loop training. It will thus be quite an issue in an open-loop design to develop an algorithm that can decode control signals for a system the brain has never controlled before, say a 12-direction rocket stabilization system.¹⁷ This places another limitation on the open-loop model, even using complex nonlinear fitting, and has driven development of the closed-loop model.

CLOSED-LOOP PERIPHERAL ARRAYS UTILIZING VISUAL FEEDBACK

In a closed-loop system that provides feedback to a control system, the differences between the two models of movement become apparent. Sensory feedback within the system is used for error correction of fine motor control, such as balance – without fine motor control and error correction for balance, the human body could not stand up. In the model with abstract cortical control, feedback corrections are processed externally to the BMI and sensory feedback to the brain is usually limited to observation. This is a subtle distinction between open-loop with visual feedback and closed-loop abstract control that includes visual feedback. In the former, the visual feedback is intended to show success of a control command being sent by the brain, whereas in the latter, visual feedback is showing success of supervisory function of a semiautonomous mechanical device.




Figure 4. Experimental Overview of Brain-Controlled Robot in a Closed-Loop With Visual Feedback Experiment. After decoding walking-related information from a monkey's brain activity while walking on a treadmill, these data were relayed from Duke University in USA to the Advanced Telecommunication Research in

¹⁷ Twelve directions are the minimum number of positive controls commands to +/- x, y, and z thrust, as well as increasing or decreasing roll, pitch, and yaw. These are six degrees of freedom but as far as a BMI control system is concerned, increasing x and decreasing x are separate commands to decode.

Japan in real time. Using visual feedback to the monkey via live streaming video, the humanoid robot in Japan was shown to execute locomotion-like movements in a similar manner as the monkey. (Reference 43)

An example of a closed loop peripheral system is the creation of the humanoid robot called CB-i (Computational Brain Interface). This system is illustrated in Figure 4. In a recent experiment, CB-i successfully mimicked the physical actions being performed by a monkey that was positioned in a remote location (Reference 43).

MEMS, ECG, AND PSOC CIRCUITRY

To address the issue of signal interface between the biological and physical systems, development has focused on hybrid devices that are created by fusing together biomolecules, cells, and other tissues, with innovative micro-size current sensors using Micro Electro Mechanical Systems (MEMS). These systems often employ the use of biomaterials and functional high polymer materials that enable the device to sense information about the human body and the environment with greater speed and sensitivity than conventional metallic sensors. These devices are built from materials and mechanisms that are compatible with the human body, and they are proving to be powerful tools for interfacing between the human body and machines.

The emergence of MEMS biotechnology has motivated various strategies to detect electrical signals generated from a stimulated region within the brain into a captured electrical response that is translated to a machine interface. One advantage of these devices is that the fusing of biological to physical circuitry is performed external to the body and the interface connection of the implants is thus biological to biological. This avoids a natural reaction of the nervous system to form scar tissue around penetrating chronic physical probes.¹⁸ Scar tissue leads to degradation of proximal signal transfer over time. These neural-tissue encapsulated chips offer significant advantages over other technologies because of their ability to be rapidly integrated into the biological environment.

Noninvasive EEG platforms often utilize a cap that covers the skull and reads the electrical signals generated from the brain on the surface of the skin atop the head. However, a technical limitation in the bandwidth of EEG-based methods often provides low information rates. Currently EEG methods are limited to 20–30 bits/min (< 0.5 bits/s). This drawback has prompted the use of a more accurate and speedy recording method based on invasive techniques such as the electrocorticogram (ECoG), where electrode arrays are placed on top of the cortex, but electrode penetration into brain tissue is minimal. ECoGs have become synonymous with the pre-surgical monitoring of epileptic seizure foci (Reference 45). Valuable insights can be gained by combining both invasive and noninvasive schemes. For example, combining functional MRI (fMRI) and intracortical recording has yielded important information about metabolic mechanisms that are most highly correlated with recorded neural activity. Microelectric neurochip developments seek to combine signal sensing and processing for bidirectional BMI systems.

Studies in the feasibility of using adaptive input–output models for reconstructing hand trajectories have recently been published. These studies have focused on the

¹⁸ Chronic devices are intended for long-duration implant, in contrast with acute implanted devices, which are utilized for a short duration and then removed. Formation of scar tissue is dependent on the material used for the electrodes.

microelectric design of chronic neural implants. In 2006, Fetz et al. illustrated an intracranial operated neural implant that could function as neural prosthetic to assist upper limb movement for individuals that suffered from spinal cord injuries. The intriguing part of the Fetz design is its ability to function as an autonomous battery operated device. This is a departure from traditional BMI devices that are cumbersome and difficult to operate. Clearly this is a step forward in the development of microwire electrode design applications (Reference 44). Figure 5 illustrates the schematic flow chart of the Fetz MEMS neurochip, also called a programmable system-on-chip (PSoC) architecture.

Programmable System on Chip Architecture




Figure 5. Schematic of the Neurochip Functional Blocks. Parallel PSoC microcontrollers record neural and muscle signals to independent memory modules. Primary PSoC also controls a constant-current stimulator circuit and communicates via IR to a PC or hand-held POA. (Reference 44)

Traditional PSoC designs operate on a high power consumption and short battery life of radiotelemetry systems (Reference 46). In the Fetz study, researchers implanted electronic devices that collected data in unrestrained monkeys. The system operated without the power limiting and cumbersome restraints exhibited by these previous systems. For this design, an onboard spike processing and a stimulator circuit were employed to allow for real-time bidirectional interface with the nervous system. After several modifications, Fetz et al. successfully created a device that was able to simultaneously record electromyogram (EMG) activity and neural activity at the implant. The lithium battery operated power source and electronic data systems were skillfully packaged within a compact percutaneous (inner skin) titanium casing that was attached to a monkey's skull; the entire implant weighed only 56 g. Neural data was acquired from 12 microwire electrodes chronically implanted in the primary motor cortex. Leads run subcutaneously from the head casing to a connector on the monkey's back. Two pairs of stainless-steel wires were inserted percutaneously into the forearm muscles where they were attached to an EMG signal recorder (Reference 44).

CHRONIC NEURAL IMPLANTS AND FMRI

While developing microelectric neural chips for functional studies appears revolutionary on the surface, it only presents a piece to the greater puzzle of understanding the feasibility between neuroprosthetic chips and BMI development. For a BMI system to succeed, researchers must understand the brain, its regions of activity and how those area correlate to real time stimulation and neural responses, and how the brain may evolve with training on use of the BMI. This will require the employment of sensitive neural mapping devices such as fMRI.




Figure 6. Implant Location. Cartoon shows location and orientation of the different electrode sites in the various layers of an animal's neocortical implant (layer thicknesses are approximately scaled). The gray band is the 200 μ m separation region between the upper and lower layers of the neocortex.

To address this question, studies performed by Parikh et al., have targeted tissue-specific layers within the motor cortex region, essentially a modified ECoG, where penetrating microelectrodes exhibit the greatest functionality in cortical prosthetic design (Reference 47). The interesting parameter of this experimental design is its ability to successfully employ a behavioral task paradigm where electrodes could accurately record brain activity and animal response based on audible and visual cues. With this information the researchers discovered that the lower layers of the cortex are more likely to encode directional information as compared to units in the upper layers. This understanding has prompted the use of prosthetic neural implants as critical instruments for developing accurate BMI models and the mapping all neurological (brain triggered) responses.

For these experiments, accurate measurement of responses was placed solely into the unique design of chronic implantable neural prosthetic devices. These devices enabled investigation of activity in the upper or lower brain tissue layers, including whether either had a preference for ipsilateral (same body side) versus contralateral (opposite side of body) movement. In addition, chronic neural implants provided an innovative method to target isolated brain regions. During this experiment a craniotomy was performed over the target cortical area. In this procedure, a 2 mm diameter hole is made into the skull to expose the dura mater. This dura matter is then removed to reveal the cortical surface. An electrode scaffold array is inserted by hand with the use of fine PTFE-coated forceps into the target cortical area. Typically, the electrode will be

localized in a manner where the top site will be below the cortical surface. The electrode assembly was then wrapped with Gel Foam (Pfizer, Inc., New York) and then cemented with dental acrylic (Reference 47). Once the experiment was concluded, precise location of the cortical arrays was obtained by passing currents through the electrodes to cause micro-lesioning, then post-mortem examination of brain slices could locate the lesions and reconstruct the precise 3-D position of the electrodes (Figure 7). This test provided an initial understanding of brain activity and the ability to utilize microelectrodes to localize specific neural regions.




Figure 7. Histology and Electrode Tracks. The left panel shows Nissl-stained coronal sections of sample slices from all animals D1–D4 showing electrode tracks or lesion marks for all seven implantations. The black line marks the boundary between the upper and lower layers. The right panel shows seven coronal sections arranged rostro-caudally, as indicated by the schematic, for one implant (D4Left) showing alternating lesions and electrode tracks which were used to reconstruct site locations. (Reference 47)

NEW ELECTRODE DESIGNS

Building upon studies like Parikh et al., designs look toward invasive systems capable of pinpointing cortical regions and monitoring their neural activity at or near single neuron resolution. Clement et al. employed microwire neural implants to combine fMRI and intracortical recordings to yield important information about metabolic mechanisms that correlate with neural activity. While modeling studies have been conducted in numerous experiments, they served as a baseline to understand the potential dangers associated with the strong static magnetic field alone or while exposed to typical MRI protocols on the surrounding tissue and the design of an MRI compatible electrical array (References 48, 49). On the surface this may seem a trivial task, however the study shows this is not only a challenging design detail, but decisive in the resulting recorded data of neural activity. Examples of the electrode designs are shown in Figure 8.




Figure 8. The Michigan Electrodes. Left: A variety of different silicon probes have been placed on the back of a U.S. penny. The copper scaffold in the background is the array of columns on the Lincoln Memorial. Right: Four 64-site probes have been assembled into a three-dimensional structure. (Reference 49)

Previous work addressed important issues related to short-term compatibility of silicon microelectrodes. In that report, identification and replacement of appropriate components of a silicon microelectrode system resulted in virtually artifact-free MR images and stable recordings of motor unit activity, although chronic studies were not performed (Reference 50). By utilizing this knowledge the authors designed a microwire electrode array capable of surviving the strong magnetic fluctuations exhibited from an fMRI. Figure 9 shows the device and its related connective components (Reference 51). Figure 10 shows the before and after MR images when connectors and structural screws compatible with MR are employed as well as the compatible electrodes.




Figure 9. Image Distortion and Custom Microwire Electrode Assembly to Improve It. (A) Image distortion induced by metallic bone screws and connectors. No feature can be identified. Spin echo sequence: TR=1s, TE=6.75ms, NT=4, Matrix=128×128. FOV=60×60mm². Experiment time=512 s. (B) (Top) Microelectrode array used in the study. (Bottom) Close-up of the connector piece that was constructed for multiple connection/disconnection cycles. Interface piece to connect the nano-miniature Omnetics connector of the head stage with the head-cap-embedded custom connector for extracellular multiunit activity monitoring. (C) Improvement in image quality after replacement of bone screws and connectors with compatible equivalents. TR=3500ms, TE=20ms, averages=2, acquisition matrix=128×96, FOV=21×21mm², slice thickness=0.4mm, total acquisition time (TA)=672sec, resolution=164μm×220μm. (Reference 51)

With this system, the authors successfully recorded spontaneous extracellular multiunit neural activity in 16 electrodes (four in each animal) for 6 weeks post-implant. Of those 16 electrodes, 12 registered data verifying distinct neural activity prior to MR exposure. To determine the overall effectiveness and feasibility of this procedure, a tissues damage assessment associated with the MRI was conducted by utilizing T₂ maps from tissue dissections (Figure 10). Thorough examination of the electrode location and resultant tissue survivability revealed little damage from the operation of the BMI. The locations of the microwires are visible as dark lines in the image (indicated by ovals).




Figure 10. Example of T₂ Variability. (A) T₂ maps from MRI rat 16 at day 7 post-implant show elevated values proximal to the implant location (circled). (B) T₂ maps from MRI rat 18 at day 30 post-implant shows no difference between the implant and control hemispheres. TR=3500ms, TE= 10, 20, 30, 40, 50, 60ms, averages=2, acquisition matrix=128×96, FOV=21×21mm², slice thickness= 0.4mm, TA=75mins, resolution=164μm×220μm. The ovals indicate the site of electrodes. (Reference 51)

The results of statistical analysis are shown in Figure 11 comparing the T_2 values obtained in the implanted hemisphere versus those in the contralateral (unimplanted) hemisphere. There was no significant difference in T_2 values between the two hemispheres. These results prove the feasibility of fMRI neural tracking and the use of microelectrode arrays. With these findings the use of microelectrical implants and fMRI will become a vital instrument in understanding neural activity in invasive BMI. With the growing use of chronic microelectric implants for neurophysiological mapping studies, future studies will be able to explore a new realm of possibilities in BMI applications.




Figure 11. T_2 Value Analysis Summary of T_2 Values in All the Image Slices That Spanned the Electrode Arrays in All Animals. There was no significant difference between the implanted and contralateral control hemispheres in the susceptibility-free regions proximal to electrodes. p-values (one-sided student t-test assuming equal variance): day0: 0.2, day7: 0.7, day14: 0.4, day28: 0.8, day42: 0.4. Combining all imaging sessions for all animals: $p=0$.

HYBRID NEURO-ROBOTIC SYSTEMS FOR TRUE CLOSED-LOOP BMI

While many innovations have been seen in BMI technologies that detect neural stimulation from direct cortical regions of monkeys and rats, another approach has been pursued through the use of lamprey eels. Chordates, like the lamprey are ideally suited for neurological studies because of their distinct and readily accessible hollow dorsal nerve cord or brain stem. In the realm of neuroscience the brain stem is often regarded as the most primitive region of the human brain. In one study, Mussa-Ivaldi et al. (References 28, 29) investigated the possibility of using the feedback from a BMI for inducing controlled plastic changes at specific synapses. Figure 12 shows the bidirectional connections between a mobile robotic device and a lamprey brain stem that has been used to investigate the repertoire of operations carried out by neurons in the reticular formation. Signals generated by the two optical sensors of the robot were translated into electrical stimuli and applied to the vestibular pathways, and to two populations of reticular neurons. The resulting discharge frequency of the reticular neurons commanded the right and left wheels of the robot. In this simple arrangement, the reticular neurons acted as a processing element that determined the closed-loop response of the neuro-robotic system to a source of light. These studies revealed that:

- Different behaviors can be generated with different electrode locations.
- The input-output relationship of the reticular synapses is well approximated by simple linear models with a recurrent dynamic component.
- The prolonged suppression of one input channel leads to altered responsiveness long after it has been restored in the variable frequency pulse generator.

The results of this lamprey study hints that cerebellar connections in mammals could provide excellent plastic interface points for full closed loop technologies.




Figure 12. A Hybrid Neuro-Robotic System. Signals from the optical sensors of a Khepera (K-team) mobile robot (bottom) are encoded by the interface into electrical stimulations, the frequency of which depends linearly on the light intensity. These stimuli are delivered by tungsten microelectrodes to the right and left vestibular pathways of a lamprey brainstem (top) that is immersed in artificial cerebrospinal fluid within a recording chamber. The electrical stimuli are delivered to the axons of the intermediate and posterior octavomotor nuclei (nOMI and nOMP, respectively). Glass microelectrodes record extracellular responses to the stimuli from the posterior rhombencephalic neurons (PRRN). Recorded signals from right and left PRRN are decoded by the interface. First, the electric artifacts generated by the stimulation impulses are removed. Then, population spikes are detected. The resulting spike train is passed through a low-pass filter, which calculates an average firing rate over a 300-millisecond firing window. The average firing rate detected from each electrode is translated into a command to the corresponding wheel of the robot. The angular velocity of the wheel is set to be proportional to this average rate.




Figure 13. Experimental Arrangements. (A): arrangement for simultaneous recordings of neuronal activity, electromyograms (EMGs), and movements in the freely behaving animal. Electrophysiological recordings and the video recordings were synchronized by pulses (1 Hz) recorded simultaneously by both systems (Synchro). (B): arrangement for recording vestibular responses in RS neurons during swimming. Central, denervated region of the body was fixed in the experimental device. Lamprey could perform locomotor-like undulatory movements by its anterior and posterior body parts (indicated by arrows). Device could be rotated manually (a, roll tilt angle).

Additional studies utilizing neurological signals generated from the lamprey have been equally promising in eliciting cognitive responses that translate a controlled response into a machine's activity. In a similar experiment, Deliagina, and coworkers (Reference 30) used the activity of reticulospinal neurons recorded from a swimming lamprey to

rotate the platform supporting the fish. The lamprey was able to stabilize the hybrid system, and this compensatory effect was most efficient in combination with undulating swimming motions. The experimental setup is seen in Figure 13. These studies demonstrate the feasibility of closed-loop interactions between a specific region of the nervous system and an artificial device. Closed-loop brain-machine interfaces offer an unparalleled opportunity to investigate how plastic changes can be guided by modulating the input signals of the neurons based on the behaviors generated by the output of the same neurons. Furthermore, in such hybrid systems it is possible to replace the neural tissue with a computational model, thus providing a direct means for testing the predictions of specific hypotheses about neural information processing. Adaptation to the variable dynamics of the limbs and of the limb environment has generated an increasing volume of research that is relevant to brain-machine interfaces.

TRIALS USING HUMAN SUBJECTS

Testing BMIs on normal human subjects presents ethical dilemma since any volunteer outside of the research team itself would likely have difficulty understanding all of the risks involved in neural implantation. Even well-known volunteers such as Jans Naumann now admit they truly did not understand the risks involved and are considering having chronic implants removed.¹⁹ Not all human trials have produced such poor results.

Significant ECoG arrays are implanted in epileptic patients prior to surgery to provide high resolution spatial localization of seizure activity (Reference 45). In these clinical tests, the arrays are implanted in the patient and then monitored until seizures takes place. This procedure can take more than a week and the patient has little to do between random onsets of seizure activity. Epilepsy is generally considered a localized pathology, almost always lateralized to the right or left lobe. The remaining brain networks, especially from the nonseizure half of the brain, are generally normal. This provides an opportunity for cohort cognitive studies since they represent minimal additional risk for volunteers undergoing the procedure as a medical necessity. Blakely et al. utilized such a patient to decode brain network patterns associated with different English phonemes (Reference 52). Mapping all of the phonemes could constitute an open-loop technology to communicate without speaking using normal language instead of learned motor patterns. Schalk and colleagues utilized a sample of five pre-surgical epileptics to demonstrate ECoG to control one- and two-dimensional cursor movement. Movement times to target were on the order of 1-2 seconds with up to 75 percent accuracy with training time on the order of 30 minutes or less (Reference 53).

Spinal cord injury (SCI) is another clinical condition where the normal brain is intact and ECoG arrays may be implanted for testing a neuroprosthetic. Researchers in Toronto recently reported good hand movement and grasping control using ECoG arrays (Reference 54). Similarly, Hochberg showed good results in 2-D cursor control, but still seconds or more slower than a mouse control, even after 90 days of training. A following task was more successful, though the accuracy for use in fine manipulations needs improvement (Reference 55).

¹⁹ Jans Naumann was the first recipient of a second-generation artificial vision system (AVS) designed by the Doherty Institute. The AVS project was based in New York, but the procedures themselves were conducted in Lisbon, Portugal, to avoid U.S. prohibitions against implant surgery.

EMG interfaces for limb amputees have advanced dramatically in the last several years, with several devices controlled by remaining muscles that have been implanted with otherwise unused nerves in an invasive process called reinnervation (References 56, 57). Twelve electrode pairs were utilized in the recording of intended arm motion. The classifier algorithm used to determine hand control was able to provide a new prediction 10 times a second, and for controls, a motion selection from 1 of 10 possibilities could be achieved in less than 170 milliseconds. Interesting with this procedure is that a direct biological amplification of the neuronal signal is achieved by connecting it to a muscle. The amplified signal is then read by a noninvasive EMG probe. Although not yet approaching normal, the speed and dexterity of these artificial limbs is impressive.

Finally, one trial utilizing a normal, healthy human self-experimenter in 2002 showed that sensation and control is possible through a peripheral array implant in the median nerve of the left arm. This trial lasted 3 months before the physical connection between the nerve and the microarray deteriorated beyond use. Subsequent examination and testing has revealed no long-term damage at the implant site (Reference 58).

OPTICAL STIMULATION OF ACTION POTENTIALS

A final invasive technology that utilizes light to gate action potentials shows promise (Reference 59). In this technology, the biological switch is controlled by a locally inserted LED and possibly guided using optical fibers, but there is no contact between the physical circuits and neurons, thus avoiding the issue of scar tissue caused by operation of invasive probes. Application to artificial vision have already interesting pre-clinical results in rat brain tissue (Reference 60).

Discussion

Any successful BMI hardware technology relies on at least one, and in most cases all, of the following three essential elements: brain plasticity through user training, neural decoding by a machine learning algorithm, and neuroscience knowledge. Advances in the latter two will help all BMI research paths, but the former element is the most research path dependent since different technologies have different training limitations such as the inherent visual feedback delay of a noninvasive BOLD monitor.

One approach to the training issue is to greatly simplify the control information bandwidth required. Research here has placed a focus on scenario-based controlling that is not from millisecond-to-millisecond updating of some 3-D trajectory like a full-duplex BMI specified in the introduction would perform, but rather is controlled by high level commands from the user in a low-bandwidth supervisory mode. It is conceivably useful to tell a wheelchair to "Go Left," "Stop," or "Veer Right," and let sensors and actuators on the wheelchair work out the trajectory details, as opposed to constantly setting and resetting the trajectory of the wheelchair. In this manner, much of the feedback error-correction is done in the physical circuit. This approach can be viewed as a series of simple asynchronous decisions, and relies on the artificial intelligence of the execution algorithms rather than tapping the general adaptability of mammalian brain architecture. Mention is made of this research path because it is important to the advancement of the field as a whole, though the concentration in this treatise is on tapping the plasticity of neural systems rather than creating a program to imitate them.

The remainder of discussion concentrates on the surveyed studies into five technology areas, and listing their advantages and disadvantages, as well as their critical barriers to further development and likely future paths of research.

NONINVASIVE ELECTRICAL DEVICES

The current state-of-the-art BMI devices that have shown the most promise in the near term utilize muscle amplification of the neural signal. This is the result of commercial investment in development. These primarily EMG based systems are noninvasive connections to existing or constructed nerve-muscle pairs. Neurally operated prosthetics have been demonstrated such as the artificial hand. Decoding algorithms are nearly as fast as muscle response itself at selecting 1 of 10 actions. Commercial products for entertainment are available for normal, healthy consumers. These entertainment devices contain few sensors but make use of more complex decoding algorithms. All sensors don't just turn on and off, rather they can relay amplitude, polarity, and frequency of muscle action, as well as detect underlying spiking signals from neurons. This combination, along with algorithms developed from a support vector machine, will allow device selection of many more actions than the simple binary combinatorics of the number of sensors. The gaming market will drive this noninvasive technology in the next 5 years with advances in signal processing from current sensors.

The EMG based systems can be scaled to a large number of sensors for a more complex control task. It has been demonstrated with amputees the feasibility to move existing nerves around to control existing muscles. There has not been a study yet to explore how many nerve-muscle pairs may be available for such use in a normal human, with or without reinnervation. Such a study could be done theoretically based on current and near term work with patients. The longer term for EMG will likely evolve toward very small, wireless chronic implants that operate near the skin surface and are easy to install and remove.

Devices based solely on noninvasive detection of EEG signals from the cortex in normal, healthy individuals are unlikely in the next 5 years. This is primarily because of muscle noise filtering issues which will need to be solved through development of better EMG sensors and detection algorithms mentioned above. Even though such devices have been shown to work in laboratory settings, moving them to the arbitrary conditions in naturalistic settings introduces significant complication to signal detection and filtering that greatly limits the information transfer bandwidth to a fraction of the target 5 bits/sec. Also an issue is the origin of surface EEG signals, which are still under theoretical study even after 80 years. Some of the frequency bands directly related to spiking activity in axons are somewhat understood, but the whole of the signals contain many transients that are likely the more interesting and possibly more robust parts of brain activity. In the medium to long term, many applications are envisioned in surface or near surface detection of cortical spiking activity or ERP, once the background separation issue in naturalistic environments is addressed. However, these signals still require localized coherent firing of tens of thousands of neurons, a biological requirement that has limited the information transfer rates of these techniques to 1 bit/sec or less, which is well below target rates but could be useful for some applications.

One can suppose an ultra-high density EEG, thousands of sensors with wireless micro probes embedded subcutaneously that track electrical activity and relative location

based on skin movement using technology similar to the operation of GPS. This system would have the same inherent problem of signal conduction through tissue from all directions. The key indicator of a future advance in EEG technology would be a study showing noninvasive spiking or ERP signal detection from a coherent group of less than 10,000 neurons. Until progress in localization on this order is demonstrated, EEG-only devices will remain low bit rate application technologies.

EEG-only sensor control of devices has been pursued with the supposition that signal transmission directly from the motor cortex will outperform conventional mechanical operation because the signal transmission delay between the neocortex and muscles in the extremities will be augmented via a physical circuit. While this is true, the inherent assumption that complex mechanical interface operation can be replaced with a 1 bit/second low-delay control pathway is incorrect for naturalistic environments. The millions of channels of two-way communication between the extremities and the brain naturally present will control naturalistic mechanical devices far better, and provides the ability to react to an ever-changing environment. Adding significant computer control and relegating the motor cortex to a supervisory role of low data bandwidth could have benefits and is a possible path to real-world improved reaction time application. Also beneficial to reaction time is the combined EMG EEG devices mentioned above since the pathways between the brain and facial muscles are far shorter than those to the extremities.

NONINVASIVE BOLD-BASED DEVICES

The detection resolution of hemodynamic responses using fMRI and NIRS type technologies is expected to improve in the near, medium, and far terms. NIRS has been deployed in naturalistic experiments, specifically, automobile driving, and it is reasonable to envision such systems becoming small enough to wear in the medium to far term. The limiting issue with BOLD technologies is that the detectable response is delayed several seconds from neuronal firing, greatly limiting the cognitive control applications available to these technologies. Progress in localization to below 0.5 mm³ in a mobile device could indicate the possibility of high bit transfer rates with the use of state vector machines, albeit with the inherent delay of the physiology. Until the localization and portability issues are addressed in tandem, BOLD-only devices are unlikely to be used for commercial cognitive control interfaces.

The vast majority of experiments in fMRI utilize the General Linear Model (GLM) and something akin to a canonical hemodynamic response. There exists significant physiological noise in these data analyses that are traditionally filtered out. The differential results from Independent Component Analysis (ICA) and GLM analysis of the same data sets indicate that there may be additional circulatory response mechanisms with faster onsets than typical BOLD. Future studies indicating that filtering or analysis methods may be ignoring important detectable signals that have inherently faster response times would provide a breakthrough path for hemodynamic response technologies. The current research to monitor on this matter is the technology known as Inverse MRI which can capture high temporal resolution (100 frames a second for single slices) T2* variations.

NONINVASIVE MAGNETIC DEVICES

Detection of magnetic fields arising from neural activity using MEG or a future technology have the same inherent limitations as EEG-only noninvasive sensors. Resolving these signals from background in any naturalistic environment will not occur without the development of a disruptive technology in shielding that is portable and lightweight enough to be worn comfortably on the head, and will attenuate background amplitude nearly 10 orders of magnitude across the entire range of neural signal frequencies, and the entire range of frequencies where the detection electronics may be sensitive. Body-temperature superconducting materials may accomplish the shielding if engineering development continues, and these could be combined with atomic magnetometer technology to produce a portable system, though the inherent limitations of the method will still be present.

INVASIVE TECHNOLOGIES – GENERAL, OPTICAL, AND EX-VIVO ENGINEERING

All invasive technologies will have to undergo substantial rigorous testing to determine the long-term side-effects of implantation and operation, as well as drift in the operational performance with extended use. Several pre-clinical devices detailed above and discussed below show promise, but the handful of human trials show that direct interface technologies based on proximal action potential stimulation and detection are still a long way off from practical application since the trials themselves in duplex operation show functional failure after only a matter of weeks. The mammalian auditory system is well studied, and the cochlear implant has been in use and continued development for more than 30 years, and there is yet to be any practical application for implanting a cochlear device into a normal healthy individual. Proximal electrode interfaces do not appear to be the answer for high bandwidth BMI, and further development in the machine-biology contact is required. If the 30-year advancement of cochlear is a realistic guide of innovation, the proximal electrical technology path to success is 30-40 years off save for a disruptive advance in the development of penetrating electrodes.

Two invasive research paths that are promising are optical stimulation and gating, and ex-vivo growth of neural cells on physical circuits.

The advantage of optical stimulation and gating is that one could theoretically design an optical link to act as a connected input dendrite on a single neuron. Theoretically, one should also be able to use spectroscopic sampling to determine if an action potential is progressing down an axon, thus providing a single neuron output (read or monitor) capability as well. This invasive but noncontact full-duplex information channel would effectively hook-up the physical computer seamlessly to the network structure of the brain. Proof of principle studies in this technology could emerge at any time, and given the demonstrated plasticity of neural networks in both closed-loop visual feedback and lamprey experiments, could provide a disruptive path for advancement of BMI.

The ex-vivo growth of neural tissue on electrode arrays provides a controlled platform for:

- Testing long-term effects of electrode use on proximal neural tissue.

- Developing complex neural networks on top of chips prior to implant.
- Development of chronic electrode materials such as biopolymers that may be suitable for implant without an ex-vivo grown sheath.
- Creating a biological to biological interface upon implant.

This technology does still rely on proximal sensing and stimulation of action potentials, but the neural tissue at the biological to physical interface is engineered rather than native, thus providing the opportunity to develop an effective bridging network to seamlessly integrate between the two systems. Development of such chronic implant capability could make feasible the nondestructive addition of cerebellar pathways for control activity that incorporates the huge potential of brain plasticity without adding significant cognitive burden on the prefrontal cortex. This would effectively turn the controlled device into part of the body as far as the brain is concerned.

IMPLANTABLE CHIPS, LADDERS AND ARRAYS

Research utilizing ECoG or neocortical ladder array technology is the state of the art in neural communication. However, stimulating nerves or peripheral pathways in this matter without damage to the receiving cells has proven difficult. Even in cochlear arrays, normal functioning of the surrounding tissue is lost and only a fraction of the quality of natural stimulation is available. Locating single neurons has also been challenging, though new electrode designs are being developed. These devices will continue to be useful for research applications such as better understanding of network structure and network function in conjunction with fMRI. BMI application will continue in the foreseeable future to be limited to instances where closed-loop with visual or other traditional sensory feedback is sufficient. Until new electrode designs develop a robust method of two-way communication with single neurons, they will not be the dominant BMI technology for use in normal, healthy individuals.

Conclusions

The gamut of modern technologies that connect neural systems to physical systems has been surveyed, with attention to the underlying physiological signals. Many examples have been presented of studies illustrating the different research paths under way. Two of these technologies show the most promise of near term and long term successful high-bandwidth integration of brain and external systems.

In the near term, noninvasive electrical sensors that primarily rely on EMG signals, but also include those weaker signals directly from neuronal firing, are expected to dominate real-world applications. The decoding of these signals from just a handful of dry sensors can produce a large combination of on-off and variable strength control commands that should surpass the information capability of traditional electro-mechanical interfaces in the next 5 years. This technology is limited, however, by the number of available EMG sites and the required coherent firing of very large numbers of neurons. To surpass these limitations, technologies capable of connections to individual neurons on a large scale will be necessary.

In the far term, an invasive approach to establish high-bandwidth duplex communication networks, 5-20 bits/second, using interfaces with single neurons

directly in the cortex or cerebellum can take the fullest advantage of neural training and processing capability. Two research paths currently showing the most promise are optical interfaces and ex-vivo preparation of electrode arrays encased in engineered neural tissue. New electrode designs that could establish two-way communication with single neurons are also promising. The cerebellar connection point is extremely interesting to perhaps provide a connection that would effectively make the external device a part of the body as far as the brain is concerned.

The current state of the art using metal electrode arrays for proximal stimulation and sensing does not appear to be advancing toward long duration viability as a commercial BMI without some unforeseen advance in two-way communication. Peripheral connections have some theoretical promise, but an example of such an implant in a human trial did not shown terrific success over several weeks it functioned. Implantation of cortical arrays in human trials has not been conducted on a wide scale for durations longer than a few weeks, and two-way electrical communication has yet to be realistically demonstrated, limiting the possible commercial application portfolio for this technology.

Many possible research directions for disruptive advances have been presented. These areas and the physiological-physical interconnection field of study as a whole should be monitored regularly for advances, as well as the ancillary technology areas of brain plasticity, neural signal decoding, and general basic neuroscience. Advances in any and all of these areas could alter the future landscape of commercial applications and dramatically alter our understanding of what is possible through thought-directed control.

¹ Alberts B. Molecular biology of the cell, 4th ed. New York: Garland Science; 2002.

² Polich J. Updating P300: an integrative theory of P3a and P3b. Clin Neurophysiol 2007 Oct;118(10):2128-48.

³ Ravden D, Polich J. On P300 measurement stability: habituation, intra-trial block variation, and ultradian rhythms. Biol Psychol 1999 Oct;51(1):59-76.

⁴ Sutton S, Braren M, Zubin J, John ER. Evoked-potential correlates of stimulus uncertainty. Science 1965 Nov 26;150(700):1187-8.

⁵ Sawaki R, Katayama J. Top-down directed attention to stimulus features and attentional allocation to bottom-up deviations. J Vis 2008;8(15):4 1-8.

⁶ Huettel SA, Song AW, McCarthy G. Functional magnetic resonance imaging. Sunderland, Mass.: Sinauer Associates, Publishers; 2004.

⁷ Kandel ER, Schwartz JH, Jessell TM. Principles of neural science. 4th ed. New York: McGraw-Hill, Health Professions Division; 2000.

⁸ Ogawa S, Tank DW, Menon R, Ellermann JM, Kim SG, Merkle H, et al. Intrinsic signal changes accompanying sensory stimulation: functional brain mapping with magnetic resonance imaging. Proc Natl Acad Sci U S A 1992 Jul 1;89(13):5951-5.

⁹ Emerging Cognitive Neuroscience and Related Technologies. Washington DC: National Research Council of the National Academies; 2008.

¹⁰ Miller MB, Van Horn JD, Wolford GL, Handy TC, Valsangkar-Smyth M, Inati S, et al. Extensive individual differences in brain activations associated with episodic retrieval are reliable over time. J Cogn Neurosci 2002 Nov 15;14(8):1200-14.

¹¹ Hatsopoulos NG, Donoghue JP. The Science of Neural Interface Systems. Annu Rev Neurosci 2009 Mar 24.

¹² Fitts PM. The information capacity of the human motor system in controlling the amplitude of movement. J Exp Psychol 1954 Jun;47(6):381-91.

¹³ Drury CG, Hoffmann ER. A model for movement time on data-entry keyboards. Ergonomics 1992;35(2):129 - 47.

¹⁴ Choi C, Micera S, Carpaneto J, Kim J. Development and quantitative performance evaluation of a noninvasive EMG computer interface. IEEE Trans Biomed Eng 2009 Jan;56(1):188-91.

¹⁵ Soukoreff RW, MacKenzie IS. Theoretical upper and lower bounds on typing speed using a stylus and soft keyboard. Behaviour & Information Technology 1995;14(6):370-9.

- ¹⁶ Cummings ML, Guerlain S. Developing operator capacity estimates for supervisory control of autonomous vehicles. *Hum Factors* 2007 Feb;49(1):1-15.
- ¹⁷ Enzinger C, Ropele S, Fazekas F, Loitfelder M, Gorani F, Seifert T, et al. Brain motor system function in a patient with complete spinal cord injury following extensive brain-computer interface training. *Exp Brain Res* 2008 Sep;190(2):215-23.
- ¹⁸ Berger H. Über das elektroencephalogramm des menschen. *Archiv für Psychiatrie und Nervenkrankheiten* 1929;87(1):527-80.
- ¹⁹ Davis P. Effects of acoustic stimuli on the waking human brain. *Journal of Neurophysiology* 1939;2:494-9.
- ²⁰ Blankertz B, Dornhege G, Krauledat M, Müller KR, Curio G. The non-invasive Berlin Brain-Computer Interface: fast acquisition of effective performance in untrained subjects. *Neuroimage* 2007 Aug 15;37(2):539-50.
- ²¹ Kubler A, Kotchoubey B, Kaiser J, Wolpaw JR, Birbaumer N. Brain-computer communication: unlocking the locked in. *Psychol Bull* 2001 May;127(3):358-75.
- ²² Bai O, Lin P, Vorbach S, Floeter MK, Hattori N, Hallett M. A high performance sensorimotor beta rhythm-based brain-computer interface associated with human natural motor behavior. *J Neural Eng* 2008 Mar;5(1):24-35.
- ²³ Iversen I, Ghanayim N, Kubler A, Neumann N, Birbaumer N, Kaiser J. Conditional associative learning examined in a paralyzed patient with amyotrophic lateral sclerosis using brain-computer interface technology. *Behav Brain Funct* 2008;4:53.
- ²⁴ Iversen IH, Ghanayim N, Kubler A, Neumann N, Birbaumer N, Kaiser J. A brain-computer interface tool to assess cognitive functions in completely paralyzed patients with amyotrophic lateral sclerosis. *Clin Neurophysiol* 2008 Oct;119(10):2214-23.
- ²⁵ Sellers EW, Kubler A, Donchin E. Brain-computer interface research at the University of South Florida Cognitive Psychophysiology Laboratory: the P300 Speller. *IEEE Trans Neural Syst Rehabil Eng* 2006 Jun;14(2):221-4.
- ²⁶ Greene K. Brain sensor for market research: A startup claims to read people's minds while they view ads. 2007 [May 12, 2009]; Available from: <http://www.technologyreview.com/Biztech/19833/?a=f>.
- ²⁷ Greene K. Connecting Your Brain to the Game Using an EEG cap, a startup hopes to change the way people interact with video games. 2007 [May 12, 2009]; Available from: <http://www.technologyreview.com/Biztech/18276/?a=f>.
- ²⁸ Popescu F, Fazli S, Badower Y, Blankertz B, Müller KR. Single trial classification of motor imagination using 6 dry EEG electrodes. *PLoS ONE* 2007;2(7):e637.
- ²⁹ van Gerven M, Jensen O. Attention modulations of posterior alpha as a control signal for two-dimensional brain-computer interfaces. *J Neurosci Methods* 2009 Apr 30;179(1):78-84.
- ³⁰ Mellinger J, Schalk G, Braun C, Preissl H, Rosenstiel W, Birbaumer N, et al. An MEG-based brain-computer interface (BCI). *Neuroimage* 2007 Jul 1;36(3):581-93.
- ³¹ Ohta H, Matsui T, Uchikawa Y. Whole-head SQUID system in a superconducting magnetic shield. *Neurol Clin Neurophysiol* 2004;2004:58.
- ³² Kraus RH, Jr., Volegov P, Matlachov A, Espy M. Toward direct neural current imaging by resonant mechanisms at ultra-low field. *Neuroimage* 2008 Jan 1;39(1):310-7.
- ³³ McDermott R, Lee S, ten Haken B, Träbesinger AH, Pines A, Clarke J. Microtesla MRI with a superconducting quantum interference device. *Proc Natl Acad Sci U S A* 2004 May 25;101(21):7857-61.
- ³⁴ Chen Y, Intes X, Taylor DR, Regatte RR, Ma H, Ntzichristos V, et al. Probing rat brain oxygenation with near-infrared spectroscopy (NIRS) and magnetic resonance imaging (MRI). *Adv Exp Med Biol* 2003;510:199-204.
- ³⁵ Luu S, Chau T. Decoding subjective preference from single-trial near-infrared spectroscopy signals. *J Neural Eng* 2009 Feb;6(1):016003.
- ³⁶ Loeb GE. Cochlear prosthetics. *Annu Rev Neurosci* 1990;13:357-71.
- ³⁷ Bailey L. New cochlear implant could improve hearing. University of Michigan News Service [serial on the Internet]. 2006; Available from: <http://www.umich.edu/news/index.html?Releases/2006/Feb06/r020606a>.
- ³⁸ Cheng YC, Brown RW, Chung YC, Duerk JL, Fujita H, Lewin JS, et al. Calculated RF electric field and temperature distributions in RF thermal ablation: comparison with gel experiments and liver imaging. *J Magn Reson Imaging* 1998 Jan-Feb;8(1):70-6.
- ³⁹ Kamitani Y, Tong F. Decoding the visual and subjective contents of the human brain. *Nat Neurosci* 2005 May;8(5):679-85.
- ⁴⁰ Kamitani Y, Tong F. Decoding seen and attended motion directions from activity in the human visual cortex. *Curr Biol* 2006 Jun 6;16(11):1096-102.
- ⁴¹ Fagg AH, Hatsopoulos NG, de Lafuente V, Moxon KA, Nemati S, Rebesco JM, et al. Biomimetic brain machine interfaces for the control of movement. *J Neurosci* 2007 Oct 31;27(44):11842-6.
- ⁴² Kim HK, Carmena JM, Biggs SJ, Hanson TL, Nicolelis MA, Srinivasan MA. The muscle activation method: an approach to impedance control of brain-machine interfaces through a musculoskeletal model of the arm. *IEEE Trans Biomed Eng* 2007 Aug;54(8):1520-9.
- ⁴³ Kawato M. Brain controlled robots. *HFSP J* 2008 Jun;2(3):136-42.
- ⁴⁴ Jackson A, Moritz CT, Mavoori J, Lucas TH, Fetz EE. The Neurochip BCI: towards a neural prosthesis for upper limb function. *IEEE Trans Neural Syst Rehabil Eng* 2006 Jun;14(2): 187-90.
- ⁴⁵ Leuthardt EC, Miller KJ, Schalk G, Rao RP, Ojemann JG. Electrocorticography-based brain computer interface--the Seattle experience. *IEEE Trans Neural Syst Rehabil Eng* 2006 Jun;14(2):194-8.
- ⁴⁶ Diorio C, Mavoori J. Computer electronics meet animal brains. *IEEE Computer* 2003;36(1):69-75.
- ⁴⁷ Parikh H, Marzullo TC, Kipke DR. Lower layers in the motor cortex are more effective targets for penetrating microelectrodes in cortical prostheses. *J Neural Eng* 2009 Apr;6(2):026004.

- ⁴⁸ Baker KB, Nyenhuis JA, Hrdlicka G, Rezai AR, Tkach JA, Shellock FG. Neurostimulation systems: assessment of magnetic field interactions associated with 1.5- and 3-Tesla MR systems. *J Magn Reson Imaging* 2005 Jan;21(1):72-7.
- ⁴⁹ Wise KD. Silicon microsystems for neuroscience and neural prostheses. *IEEE Eng Med Biol Mag* 2005 Sep-Oct;24(5):22-9.
- ⁵⁰ Martinez Santiesteban FM, Swanson SD, Noll DC, Anderson DJ. Magnetic resonance compatibility of multichannel silicon microelectrode systems for neural recording and stimulation: design criteria, tests, and recommendations. *IEEE Trans Biomed Eng* 2006 Mar;53(3):547-58.
- ⁵¹ Paralikar KJ, Neuberger T, Matsui JT, Barber AJ, Webb A, Clement RS. Feasibility and safety of longitudinal magnetic resonance imaging in a rodent model with intracortical microwire implants. *J Neural Eng* 2009 Apr 15;6(3):34001.
- ⁵² Blakely T, Miller KJ, Rao RP, Holmes MD, Ojemann JG. Localization and classification of phonemes using high spatial resolution electrocorticography (ECoG) grids. *Conf Proc IEEE Eng Med Biol Soc* 2008;2008:4964-7.
- ⁵³ Schalk G, Miller KJ, Anderson NR, Wilson JA, Smyth MD, Ojemann JG, et al. Two-dimensional movement control using electrocorticographic signals in humans. *J Neural Eng* 2008 Mar;5(1):75-84.
- ⁵⁴ Marquez-Chin C, Popovic MR, Cameron T, Lozano AM, Chen R. Control of a neuroprosthesis for grasping using off-line classification of electrocorticographic signals: case study. *Spinal Cord* 2009 Apr 21.
- ⁵⁵ Hochberg LR, Serruya MD, Friebs GM, Mukand JA, Saleh M, Caplan AH, et al. Neuronal ensemble control of prosthetic devices by a human with tetraplegia. *Nature* 2006 Jul 13;442(7099):164-71.
- ⁵⁶ Kuiken TA, Li G, Lock BA, Lipschutz RD, Miller LA, Stubblefield KA, et al. Targeted muscle reinnervation for real-time myoelectric control of multifunction artificial arms. *JAMA* 2009 Feb 11;301(6):619-28.
- ⁵⁷ Kuiken T. Targeted reinnervation for improved prosthetic function. *Phys Med Rehabil Clin N Am* 2006 Feb;17(1):1-13.
- ⁵⁸ Warwick K, Gasson M, Hutt B, Goodhew I, Kyberd P, Andrews B, et al. The application of implant technology for cybernetic systems. *Arch Neurol* 2003 Oct;60(10):1369-73.
- ⁵⁹ Nagel G, Szellas T, Huhn W, Kateriya S, Adeishvili N, Berthold P, et al. Channelrhodopsin-2, a directly light-gated cation-selective membrane channel. *Proc Natl Acad Sci U S A* 2003 Nov 25;100(24):13940-5.
- ⁶⁰ Degenaar P, Grossman N, Memon MA, Burrone J, Dawson M, Drakakis E, et al. Optobionic vision--a new genetically enhanced light on retinal prosthesis. *J Neural Eng* 2009 Jun;6(3):035007.